

Certificate



Certificate no.

TA 50012689 01

Applicant:

Mean Well Enterprises Co., Ltd.
4Fl.-2, No. 26, Wu-Chuan 2nd Road
HSIN CHUANG, TAIPEI HSIEN 242
TAIWAN, R.O.C.

Manufacturer:

Mean Well Enterprises Co., Ltd.
4Fl.-2, No. 26, Wu-Chuan 2nd Road
HSIN CHUANG, TAIPEI HSIEN 242
TAIWAN, R.O.C.

Test report no.: ZTW1-PYH 10002587 001

Client Reference: 90-0877-1/PSE

Tested to:

EN 60601-1:1990+A1+A2
IEC 60601-1:1988+A1+A2

Test object: (Open Framed Switching Power Supply)

License Fee - Unit

Type Designation	: MPS-30-x	8
x stands for	: 5, 12, 15, 24, 27 or 48	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.8A	
max. Ambient Temperature:	45°C	
Protection Class	: I	


Continued on page 02

Appendix: 1



9

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue
(day/mo/yr)

30.04.2002

Dipl.-Ing. R. Scheller

Dipl.-Ing. A. Klinker

Certificate



TÜV

Group Asia

Certificate no.

TA 50012689 02

Applicant:

Mean Well Enterprises Co., Ltd.
4Fl.-2, No. 26, Wu-Chuan 2nd Road
HSIN CHUANG, TAIPEI HSIEN 242
TAIWAN, R.O.C.

Manufacturer:

Mean Well Enterprises Co., Ltd.
4Fl.-2, No. 26, Wu-Chuan 2nd Road
HSIN CHUANG, TAIPEI HSIEN 242
TAIWAN, R.O.C.

Test report no.: ZTW1-PYH 10002587 001

Client Reference: 90-0877-1/PSE

Tested to:

EN 60601-1:1990+A1+A2
IEC 60601-1:1988+A1+A2

Test object: (Open Framed Switching Power Supply)

License Fee - Unit

as page 01

Continuation

Type Designation

MPS-30-5

MPS-30-12

MPS-30-15

MPS-30-24

MPS-30-27

MPS-30-48

Remarks: Primary and secondary circuits are separated according to method 4 of clause 17.g.

Appendix: 1

Output

DC 5V/5.0A

DC 12V/2.5A

DC 15V/2.0A


DC 24V/1.2A

DC 27V/1.1A

DC 48V/0.6A



Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue

(day/mo/yr)

30.04.2002

Dipl.-Ing. R. Scheller

Dipl.-Ing. A. Klinker

Certificate



Certificate no.

TA 50012689 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Test report no.: ZTW1-BHU 10002587 002

Client Reference: 94-0965/PSE

Tested to:

EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2

Certified Product: (Open Framed Switching Power Supply)

License Fee - Units

as page 01

Change

License Holder: see above


Factory : see above

Test Requirement: see above

Appendix: 1



Licensed Test mark:





TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue
(day/mo/yr)

27/09/2005


Dipl.-Ing. R. Scheller


Dipl.-Ing. A. Klinker