

Certificate



TÜV

Group Asia

Certificate no.

TA 50009713 01

Applicant:

Mean Well Enterprises Co., Ltd.
4Fl.-2, No. 26, Wu-Chuan 2nd Road
HSIN CHUANG, TAIPEI HSIEN 242
TAIWAN, R.O.C.

Manufacturer:

Mean Well Enterprises Co., Ltd.
4Fl.-2, No. 26, Wu-Chuan 2nd Road
HSIN CHUANG, TAIPEI HSIEN 242
TAIWAN, R.O.C.

Test report no.: ZTW1-PYH 10001834 001

Client Reference: 12006346

Tested to:

EN 60601-1:1990+A1+A2
IEC 60601-1:1988+A1+A2

Test object: Switching Power Supply

License Fee - Unit

Type Designation: MPx-yz
x stands for : S, D or T
y stands for : 45 or 65
z stands for : -3.3, -5, -7.5, -12, -13.5, -15,
-24, -27, -48, A, B or C
Rated Voltage : AC 100-240V, 50/60Hz
Rated Current : a) 1.2A (for MPx-45z)
b) 1.6A (for MPx-65z)
Output Voltage : see Constructional Data Form
Output Current : see Constructional Data Form
Protection Class: I

6
1
1
1

Appendix: 1



9

Licensed Test mark:

TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue
(day/mo/yr)

31.01.2002



Dipl.-Ing. R. Scheller

Dipl.-Ing. A. Klinker

Certificate



Group Asia

Certificate no.

TA 50009713 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Test report no.: ZTW1-BHU 10001834 002

Client Reference: 93-0488/PSE

Tested to:

EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2

Certified Product: Switching Power Supply

License Fee - Units

as page 01

Change

License Holder : see above


Factory : see above

Test Requirement: see above

Appendix: 1



Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue

(day/mo/yr)

29/06/2004

i. v. d. J.
Dipl.-Ing. R. Scheller

i. d. w. K.
Dipl.-Ing. A. Klinker

TÜV Rheinland Taiwan Ltd. - Tel (02) 2516-6040 - Fax (02) 2504-5040

Certificate



Certificate no.

TA 50009713 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-SCL 10001834 003

Client Reference: 8318801255/CC140910

Tested to: IEC 60601-1:1988+A1+A2
EN 60601-1:1990+A1+A2+A13

Certified Product: Switching Power Supply

License Fee - Units

as page 01


Addition

Factory : see above

Remark: This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

30/09/2010